

1. Crist B, Spisak D. Evaluation of Induced Variance of Physical Parameters on the Calibrated USP Dissolution Apparatus 1 and 2. [Dissolution Tech 2005;12:28-31](#).
2. Kukura J, Baxter JL, Muzzio FJ. Shear distribution and variability in the USP Apparatus 2 under turbulent conditions. [Int J Pharm 2004;26:9-17](#).
3. Qureshi SA. A new crescent-shaped spindle for drug dissolution testing - but why a new spindle? [Dissolution Technol 2004;11:13-18](#).
4. Qureshi SA. Improved drug dissolution and product characterization using the crescent-shaped spindle. [J Pharm Pharmacol 2004;56:1135-1141](#).
5. Qureshi, SA. Choice of rotation speed (rpm) for bio-relevant drug dissolution testing using a crescent-shape spindle. [Eur J Pharm Sci. 2004;23:271-275](#).
6. Qureshi SA, Shabnam J. Applications of a new device (spindle) for improved characterization of drug release (dissolution) of pharmaceutical products. [Eur J Pharm Sci 2003;22:291-297](#).
7. Healy AM, McCarthy LG, Gallagher KM *et al*. Sensitivity of dissolution rate to location in the paddle dissolution apparatus. [J Pharm Pharmacol. 2002;54:441-4](#).
8. Qureshi SA, Shabnam J. Cause of high variability in drug dissolution testing and its impact on setting tolerances. [Eur J Pharm Sci 2001;12:271-6](#).
9. Qureshi SA, McGilveray IJ. Typical variability in drug dissolution testing: Study with USP and FDA calibrator tablets and a marketed drug (glibenclamide) product. [Eur J Pharm Sci 1999;7:249-258](#).
10. Meyer MC, Straughn AB, Mhatre RM *et al*. Lack of in vivo/in vitro correlations for 50 mg and 250 mg primidone tablets. [Pharm Res 1998;15:1085-9](#).
11. Achanta AS, Gray VA, Cecil TL *et al*. Evaluation of the performance of prednisone and salicylic acid calibrators. [Drug Develop Ind Pharm 1995;21:1171-1182](#).
12. Cox DC, Furman WB, Page DP. Systematic error associated with apparatus 2 of the USP dissolution test IV: effect of air dissolved in the dissolution medium. [J Pharm Sci 1983;9:1061-4](#).
13. Cox DC, Furman WB, Thornton LK, Moore TW, Jefferson EH. Systematic error associated with apparatus 2 of the USP dissolution test III: limitations of calibrators and the USP suitability test. [J Pharm Sci. 1983;72\(8\):910-3](#).