

USP Dissolution Performance Verification Test (PVT) limits Changed (again!). Apparently retrospectively!

Recent ([link](#))

Table 1. Performance Verification Test (PVT) limits (values apply only to Lot R031Y0)

Apparatus	# of vessels	Single-Stage		Two-Stage							
				1 st Stage of Two Stages		2 nd Stage of Two Stages					
		GM*	%CV	GM*	%CV	GM*	%CV				
1	6	51-80	11	55-74	8.2	51-80	11				
	7										
	8										
	12							na			
	14							na			
2	6	27-41	6.4	29-38	4.7	27-41	6.3				
	7		6.3								
	8		6.2					4.8	6.2		
	12		6.3	na							
	14		6.2	na							
	* Percent of the labeled amount of prednisone dissolved at 30 minutes at 50-rpm										

Copyright 2014 The United States Pharmacopeial Convention. All rights reserved.
 USP Certificate
 Page 2 of 5

Certificate Date: 03Nov2015
 USP Template No. CERT1_4-02
 Effective June 05, 2014

Previous ([link](#))

Table 1. Performance Verification Test (PVT) limits (values apply only to Lot R001B1)

Apparatus	# of vessels	Single-Stage		Two-Stage			
				1 st Stage of Two Stages		2 nd Stage of Two Stages	
		GM*	%CV	GM*	%CV	GM*	%CV
1	6	55-82	12	59-77	8.8	55-82	12
	7		11				11
	8		na				
	12	54-83	12	na			
	14	55-83	na				
2	6	26-38	6.6	28-37	4.9	26-38	6.4
	7		6.4				6.2
	8		6.3				6.1
	12	26-39	6.5	na			
	14	26-39	6.4	na			
* Percent of the labeled amount of prednisone dissolved at 30 minutes at 50-rpm							

Copyright 2014 The United States Pharmacopeial Convention. All rights reserved.
 USP Certificate
 Page 2 of 4

Certificate Date: 13July2015 Version 01
 USP Template No. CERT1_4-02
 Effective June 05, 2014